

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 30 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130129</u>
Name <u>JAMES MICHELS</u>	ID# <u>6502</u>	Date <u>4/27/10</u>
A Agency <u>USB POLICE DEPT</u>		Phone # <u>852-6111</u>
Instrument Location <u>BARROW - SQUAD ROOM</u>		
B Alco S/N <u>X301568</u> Target Value <u>.084</u> High Pressure <u>800 psi</u>		
Alco Test Values <u>.085</u> <u>.083</u> 1 st Alco 2 nd Alco		
Signature <u>[Signature]</u> <u>SGT. MICHELS</u>		
(OVER)		

BMB
5/3/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of June, 2010.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

APRIL 27, 2010

OPERATOR'S NAME:

NICHOLS/JAMES/D

OPERATOR'S NUMBER: 6502

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: DMW

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .084

ALCO S/N: X301568

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.06 in

ALCO TARGET	.084	22:32
BLANK TEST	.000	22:33
INTERNAL STANDARD	VERIFIED	22:33
ALCO TV 30.04 in	.085	22:34
BLANK TEST	.000	22:35
SUBJECT SAMPLE	.000	22:35
BLANK TEST	.000	22:36
ALCO TV 30.06 in	.083	22:36
BLANK TEST	.000	22:37

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

APRIL 27, 2010

TIME 22:38

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	40c
BAROMETER:	30.06 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~"